

**LABEL IN PART:** (Box) "Cardinal Fever Thermometer Kind—Oral."

**ACCOMPANYING LABELING:** Leaflets designated "Certificate of Accuracy."

**RESULTS OF INVESTIGATION:** Examination revealed that 8 out of 24 thermometers taken from this lot failed to comply with the requirement for accuracy specified in Commercial Standard CS1-52 issued by the National Bureau of Standards of the Department of Commerce when tested as described in Commercial Standard CS1-52.

**LIBELED:** 2-7-56, N. Dist. Ohio.

**CHARGE:** 501 (c)—the quality of the article, when shipped, fell below that which it purported and was represented to possess; and 502 (a)—the statement in the labeling accompanying the article when shipped, namely, "We, the undersigned manufacturers, hereby certify that this registering clinical thermometer has been tested and found to meet all the requirements and tests specified in Commercial Standard CS1-52, as developed by the trade under the procedure of the Commodity Standards Division, and issued by the United States Department of Commerce," was false and misleading.

**DISPOSITION:** 3-7-56. Default—destruction.

**5017. Clinical thermometers.** (F. D. C. No. 38921. S. No. 47-488 M.)

**QUANTITY:** 312 *clinical thermometers* at Newark, N. J.

**SHIPPED:** 12-10-55, from Brooklyn, N. Y., by Cardinal Thermometer Co.

**LABEL IN PART:** (Box) "Cardinal Fever Thermometer Kind—Oral."

**ACCOMPANYING LABELING:** Leaflets designated "Certificate of Accuracy."

**RESULTS OF INVESTIGATION:** Examination revealed that 3 out of 24 thermometers taken from this lot failed to comply with the requirement for accuracy specified in Commercial Standard CS1-52 issued by the National Bureau of Standards of the Department of Commerce when tested as described in Commercial Standard CS1-52.

**LIBELED:** 1-24-56, Dist. N. J.

**CHARGE:** 501 (c)—the quality of the article, when shipped, fell below that which it purported and was represented to possess; and 502 (a)—the statement in the labeling accompanying the article when shipped, namely, "We, the undersigned manufacturers, hereby certify that this registering clinical thermometer has been tested and found to meet all the requirements and tests specified in Commercial Standard CS1-52, as developed by the trade under the procedure of the Commodity Standards Division, and issued by the United States Department of Commerce," was false and misleading.

**DISPOSITION:** 3-5-56. Default—destruction.

**5018. Clinical thermometers.** (F. D. C. No. 38720. S. No. 29-942 M.)

**QUANTITY:** 176 *clinical thermometers* in individual cartons at Paterson, N. J.

**SHIPPED:** Between 12-4-53 and 9-9-55, from Brooklyn, N. Y., by Cornell Instrument Co.

**LABEL IN PART:** (Carton) "Cornell Fever Thermometer Baby Special."

**ACCOMPANYING LABELING:** Inserts designated "Certificate of Examination."

**RESULTS OF INVESTIGATION:** Examination showed that 6 of 24 thermometers taken from this lot failed to meet the labeled standard of accuracy.

**LIBELED:** 11-28-55, Dist. N. J.